

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 535 050 BL

Volumetric pipettes, 5 ml, blue grad., AS, conf. cert.


Batch: 30-18

Serial:

Measuring Nr. 23526

Batch size: 1200

Volume:	5,0000 ml
Tolerance:	$\pm 0,0150$ ml
Average value:	5,0018 ml
Deviation:	0,0025 ml
Precision (CV):	0,0507 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator: 9

Date: 22.08.2018

Balance: Sartorius -
LE244S-OCE

DE-M