

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 055

**Volumetric pipettes, 5,5 ml, AS, conf. cert.**


**Batch:** 10-26

**Serial:**

**Measuring Nr.** 77027

**Batch size:** 50

<b>Volume:</b>	<b>5,5000 ml</b>
<b>Tolerance:</b>	$\pm 0,0150$ ml
<b>Average value:</b>	5,4961 ml
<b>Deviation:</b>	0,0016 ml
<b>Precision (CV):</b>	0,0288 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

<b>Operator:</b>	7
<b>Date:</b>	18.03.2026
<b>Balance:</b>	Sartorius - LE244S-OCE

DE-M