

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 060

**Volumetric pipettes, 6 ml, AS, KB**

**Batch:** 49-17


**Serial:**

**Measuring Nr.** 16485

**Batch size:**

500

<b>Volume:</b>	<b>6,0000 ml</b>
<b>Tolerance:</b>	$\pm 0,0150$ ml
<b>Average value:</b>	6,0039 ml
<b>Deviation:</b>	0,0019 ml
<b>Precision (CV):</b>	0,0323 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

14.12.2017

**Balance:**

Sartorius -  
LE244S-OCE

DE-M