

# Certificate of performance



ISO 9001  
CERTIFIED

**Article:** 3 535 920

**Volumetric pipettes, 200 ml, AS, conf. cert.**

**Batch:** 46-24


**Serial:**

**Measuring Nr.** 70689

**Batch size:**

19

<b>Volume:</b>	200,000 ml
<b>Tolerance:</b>	$\pm 0,0800$ ml
<b>Average value:</b>	199,9746 ml
<b>Deviation:</b>	0,0188 ml
<b>Precision (CV):</b>	0,0094 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

**witeg Labortechnik GmbH**

Am Bildacker 16  
97877 Wertheim  
Germany  
+49 (0) 9342 / 9301-0  
info@witeg.de

**Operator:**

4

**Date:**

27.11.2024

**Balance:**

Sartorius - LA  
1200 S

DE-M