

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 555 050

Volumetric pipettes, 5 ml, 2 marks, conf. cert.


Batch: 39-23

Serial:

Measuring Nr. 64703

Batch size: 499

Volume:	5,0000 ml
Tolerance:	$\pm 0,0150$ ml
Average value:	5,0016 ml
Deviation:	0,0028 ml
Precision (CV):	0,0562 %



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	7
Date:	19.10.2023
Balance:	Sartorius - LE244S-OCE

DE-M