

Certificate of performance



ISO 9001
CERTIFIED

Article: 3 555 100

Volumetric pipettes, 10 ml, with 2 marks, AS, conf. cert.

Batch: 48-17


Serial:

Measuring Nr. 17657

Batch size:

1

Volume:	10,000 ml
Tolerance:	$\pm 0,0200$ ml
Measured value:	10,0061 ml



These final test values refer to the delivered nominal volume (max. volume), instrument and liquid (dist. water) at equilibrium with ambient temperature (20 °C). The test was effected according to ISO 648

witeg Labortechnik GmbH

Am Bildacker 16
97877 Wertheim
Germany
+49 (0) 9342 / 9301-0
info@witeg.de

Operator:	9
Date:	29.01.2018
Balance:	Sartorius - LE244S-OCE

DE-M